

American National Standard



ANSI/AAMI/ IEC 60601- 1-8:2006 & A1:2012

MEDICAL ELECTRICAL EQUIPMENT – Part
1-8: General requirements for basic safety
and essential performance – Collateral
Standard: General requirements, tests and
guidance for alarm systems in medical
electrical equipment and medical electrical
systems

Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

**MEDICAL ELECTRICAL EQUIPMENT – Part 1-8:
General requirements for basic safety and essential
performance – Collateral Standard: General
requirements, tests and guidance for alarm systems in
medical electrical equipment and medical electrical
systems**

Approved 18 October 2013 by
Association for the Advancement of Medical Instrumentation

Approved 21 October 2013 by
American National Standards Institute, Inc.

Abstract: This standard defines the roles, responsibilities and activities that are necessary for risk management of IT-networks incorporating medical devices to address safety, effectiveness, and data and system security.

Keywords: medical device, risk management, information technology, interoperability

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI, or by visiting the AAMI website at www.aami.org.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of IEC, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: 703-525-4890; Fax: 703-525-1067.

Printed in the United States of America

ISBN 1-57020-517-5

This is a preview. Click here to purchase the full publication.

CONTENTS

Glossary of equivalent standards.....	v
Committee representation.....	vi
Background of ANSI/AAMI adoption of IEC 60601-1-8:2006 & A1:2012	viii
FOREWORD	ix
INTRODUCTION TO THE AMENDMENT	7
INTRODUCTION	xii
1 * Scope, object and related standards	1
1.1 Scope	1
1.2 Object	1
1.3 Related standards	1
2 Normative references	2
3 Terms and definitions	3
4 General requirements	7
5 ME EQUIPMENT identification marking and documents	7
5.1 Indicator lights and controls	7
5.2 ACCOMPANYING DOCUMENTS	7
6 ALARM SYSTEMS	8
6.1 ALARM CONDITION	8
6.2 * Disclosures for INTELLIGENT ALARM SYSTEM	9
6.3 Generation of ALARM SIGNALS	10
6.4 * Disclosure of delays	17
6.5 ALARM PRESETS	18
6.6 ALARM LIMIT	21
6.7 * ALARM SYSTEM security	22
6.8 * ALARM SIGNAL inactivation states	22
6.9 * ALARM RESET	25
6.10 * NON-LATCHING and LATCHING ALARM SIGNALS	25
6.11 * DISTRIBUTED ALARM SYSTEM	26
6.12 * ALARM CONDITION SYSTEM logging	28
Annex A (informative) General guidance and rationale	30
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	69
Annex C (normative) Symbols on marking	73
ANNEX D (informative) Guidance for auditory ALARM SIGNALS	79
ANNEX E (informative) Verbal ALARM SIGNALS	81
ANNEX F (normative) * Reserved melodies for ALARM SIGNALS	83
Bibliography	84

Index of defined terms used in this collateral standard.....	86
Figure 1 – Illustration of temporal characteristics of auditory ALARM SIGNALS	14
Figure A.1 – Graphical representation of components of ALARM SYSTEM delay.....	50
Table 1 – ALARM CONDITION priorities Determination of ALARM CONDITIONS and assignment of priorities	9
Table 2 – Characteristics of alarm indicator lights.....	11
Table 3 – * Characteristics of the BURST of auditory ALARM SIGNALS	13
Table 4 – * Characteristics of the PULSE of auditory ALARM SIGNALS	14
Table 5 – ALARM SIGNAL inactivation states	25
Table A.1 – Reference interpretation of Table F.1	68
Table A.2 – Reference interpretation of Table F.2	68
Table B.1 – Cross-reference of marking	69
Table B.2 – Cross-reference of ACCOMPANYING DOCUMENTS.....	70
Table B.3 – Cross-reference of instructions for use	71
Table B.4 – Cross-reference of technical description	72
Table C.1 – Graphical symbols for ALARM SYSTEMS.....	73
Table C.2 – Alternative ALARM SYSTEM related markings	78
Table D.1 – Attributes of perceived urgency	79
Table F.1 – * Equipment encoded auditory ALARM SIGNALS categorized by ALARM CONDITION and priority complying with Table 3 and Table 4	83
Table F.2 – * Auditory LOW PRIORITY ALARM SIGNAL complying with Table 3 and Table 4	83

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf

Committee representation

Association for the Advancement of Medical Instrumentation

Information Technology Networks Incorporating Medical Devices Committee

The adoption of IEC 60601-1-8 as an American National Standard was initiated by the AAMI Medical Device Alarms Committee. The AAMI Medical Device Alarms Committee also functions as a U.S. Technical Advisory Group to the relevant work in the International Electrotechnical Commission (IEC). U.S. representatives from the AAMI Medical Device Alarms Committee (U.S. Sub-TAG for IEC/SC62A/JWG 2-ISO/TC 121/SC 3) played an active part in developing the IEC standard.

Committee approval of this document does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Medical Device Alarms Committee** had the following members:

<i>Cochairs</i>	Frank E Block, Jr., MD David G. Osborn, Philips Electronics North America
<i>Members</i>	Pat Anglin-Regal, Massachusetts General Hospital Emily Arnould, RN BSN, Hospira Worldwide Inc. Shashi Avadhani, CBET CCE MBA, Crothall Clinical Equipment Services Nancy Blake, RN, Children's Hospital Los Angeles Patricia Bourie, RN MS, Beth Israel Deaconess Medical Center J. Tobey Clark, CCE, University of Vermont Christopher S. Connelly, Miami Valley Hospital Todd Cooper, West Health Institute Randolph J. Cremer, CBET, Deborah Heart & Lung Center Conor Curtin, Fresenius Medical Care Renal Therapies Group David DeBelser, Smiths Medical James E. Eberhart, BSEE, Meritus Medical Center Judy Edworthy, BA PhD, University of Plymouth School of Psychology Jane Foley, RN, Beth Israel Deaconess Medical Center Shawn Forrest, FDA/CDRH Marjorie Funk, Yale University School of Nursing Daryle Jean Gardner-Bonneau, PhD, Bonneau and Associates Karen Giuliano, Fluidnet Corporation Rodolfo I. Godinez, MD, PhD James Greenberg, Cincinnati Children's Hospital Medical Center Brian Gross, Philips Electronics North America David Hengl, Draeger Medical Systems Inc. David H. Hoffmeister, Baxter Healthcare Corporation Dean A. Hooper, HE Consulting Elizabeth Howard, DaVita Total Renal Care Inc. Xiao Hu, PhD, UCLA Chantal Kangudie, Sunnybrook Health Sciences Centre Ted J. Klefisch, Covidien Todd Konieczny, Intertek Colleen Lindell, RN, Regions Hospital Alan Lipschultz, CCE PE CSP, HealthCare Technology Consulting LLC Shawn O'Connell, MS, RN, B Braun of America Inc. Yashaswini Patwardhan, Baxter Healthcare Corporation Melanie Quinton, Kaiser Permanente (US) Cadathur Rajagopalan, PhD SMIEEE, Mindray DS USA Inc. Sue E. Sendelbach, PhD RN, Abbott Northwestern Hospital Elena Simoncini, VA New England Healthcare System Bob Steurer, Spacelabs Medical Inc. Katherine Tindall, RN, Welch Allyn Inc. Sandra Tobar, RN, St John Providence Health System Stephen Treacy, GE Healthcare

Steven A. White, NxStage Medical Inc.
Daidi Zhong, PhD, Chongqing University

Alternates

Robert J. Alberte, Jr., GE Healthcare
Keith B. Anderson, BSEE, Smiths Medical
Martin Crnkovich, Fresenius Medical Care Renal Therapies Group
Julian M. Goldman, MD, Massachusetts General Hospital
Randy Good, Covidien
Thomas M. Judd, MS PE CCE CPHQ FACCE, Kaiser Permanente (US)
Joshua Kim, Welch Allyn Inc.
Gary N. Mills, Hospira Worldwide Inc.
Yashaswini Patwardhan, Baxter Healthcare Corporation
Linda Ricci, FDA/CDRH

NOTE--Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Background of ANSI/AAMI adoption of IEC 60601-1-8:2006 and IEC 60601-1-8:2006/A1:2012

As indicated in the foreword to the main body of this document (page ix), the International Electrotechnical Commission (IEC) is a worldwide federation of national standards bodies. The United States is one of the IEC members that took an active role in the development of this standard and its amendment, which was developed by IEC Technical Committee 62, Electrical equipment in medical practice, Subcommittee (SC) 62A, Common aspects of electrical equipment used in medical practice, to provide terminology, requirements, general recommendations and guidance for medical electrical equipment and medical electrical systems manufacturers and for technical committees responsible for particular standards.

U.S. participation in IEC/SC 62A is organized through the U.S. Technical Advisory Group to IEC/SC 62A, administered by AdvaMed. Experts from the United States made a considerable contribution to this standard.

ANSI/AAMI/IEC 60601-1-8:2013 was approved by the American National Standards Institute (ANSI) on 21 October 2013.

AAMI and ANSI procedures require that standards be reviewed every five years and, if necessary, revised to reflect technological advances that may have occurred since publication.

AAMI (and ANSI) have adopted other ISO standards. See the Glossary of Equivalent Standards for a list of ISO standards adopted by AAMI, which gives the corresponding U.S. designation and the level of equivalency with the ISO standard.

As used within the context of this document, “shall” indicates requirements strictly to be followed to conform to the recommended practice. “Should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the recommended practice. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

AAMI (and ANSI) have adopted other IEC and ISO standards. See the Glossary of Equivalent Standards for a list of IEC and ISO standards adopted by AAMI, which gives the corresponding U.S. designation and the level of equivalency with the IEC and ISO standard.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data comes to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 4301 N. Fairfax Dr. Suite 301, Arlington, VA 22203-1633.

NOTE—Beginning with the foreword on page ix, this American National Standard is identical to IEC 60601-1-8:2006 and IEC 60601-1-8:2006/A1:2012.

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-8: General requirements for basic safety and essential performance –

Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of IEC 60601-1-8 consists of the second edition (2006) [documents 62A/519/CDV and 62A/537A/RVC] and its amendment 1 (2012) [documents 62A/824/FDIS and 62A/837/RVD]. It bears the edition number 2.1.

The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.