


American National Standard



ANSI/AAMI/ IEC 62304: 2006 & A1:2016 (Consolidated Text)

Medical device software—
Software life cycle processes

Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

Medical device software—Software life cycle processes

Approved 18 December 2015 by
AAMI

Approved 7 April 2016 by
American National Standards Institute, Inc.

Abstract: This standard applies to the development and maintenance of MEDICAL DEVICE SOFTWARE when software is itself a MEDICAL DEVICE or when software is an embedded or integral part of the final MEDICAL DEVICE. This standard describes PROCESSES that are intended to be applied to software which executes on a processor or which is executed by other software (for example an interpreter) which executes on a processor.

Keywords: medical device, software, life cycle

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are voluntary, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

AAMI
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2016 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 276-0793.

Printed in the United States of America

ISBN 1-57020-598-1

This is a preview. Click here to purchase the full publication.

Contents	Page
Glossary of equivalent standards	iv
Committee representation	v
Foreword	vii
Introduction.....	ix
Introduction to Amendment 1	xii
1 Scope.....	1
2 * Normative references.....	2
3 * Terms and definitions	2
4 * General requirements	7
5 Software development PROCESS.....	10
6 Software maintenance PROCESS	19
7 * Software RISK MANAGEMENT PROCESS.....	20
8 * Software configuration management PROCESS.....	22
9 * Software problem resolution PROCESS	23
Annex A (informative) Rationale for the requirements of this standard	25
Annex B (informative) Guidance on the provisions of this standard.....	28
Annex C (informative) Relationship to other standards.....	43
Annex D (informative) Implementation	62
Bibliography.....	64
Index of defined terms	66

Glossary of equivalent standards

International Standards or Technical Reports adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf

Committee representation

Association for the Advancement of Medical Instrumentation Software Work Group

The publication of ANSI/AAMI/IEC 62304:2006/A1 as a new American National Standard was initiated by the AAMI Information Technology Networks Work Group, which also functions as a U.S. Technical Advisory Group to the relevant work in the International Organization for Standardization (ISO) ISO/TC210 – IEC/SC62A JWG2. U.S. representatives from the AAMI Software Work Group participate as US experts on the ISO committee.

At the time this document was published, the **AAMI Software Work Group** had the following members:

Cochairs: Sherman Eagles
John Murray

Members: Mark Adams, Steris Corporation
Michael Attili, Amaxo Inc
Michael Brendel, Spacelabs Healthcare
Frank Clay
Kimberly Colasanti, Welch Allyn Inc
Todd Cooper, Center for Medical Interoperability
Conor Curtin, Fresenius Medical Care
Richard DeLaCruz, Silver Lake Group Inc
Theresa Dennis, Sterigenics International
Harsh Dharwad, Hospira Worldwide Inc
Sherman Eagles, SoftwareCPR
Plamena Entcheva-Dimitrov, Regulatory Consultant
Christie Evans, Hill-Rom Holdings
Chris Flahive, Chris Flahive Associates
Rick Hampton, Premier Inc - Charlotte, NC
Ed Heierman, Abbott Laboratories
Jeremy Jensen, Boston Scientific Corporation
Clint Johnson, UC Davis Medical Cntr
Rob Johnson, Baxter Healthcare Corporation
Michelle Jump, Stryker Instruments Division
Jim Kainec, Steris Corporation
Patty Krantz-Zuppan, Medtronic Inc WHQ Campus
Alan Kusinitz, SoftwareCPR
Yimin Li, St Jude Medical Inc
Jared Mauldin, Integrated Medical Systems
Mary Beth McDonald, Mary Beth McDonald Consulting
Mulugeta Mideksa
Mark Miller, Covidien
John Murray, FDA/CDRH
Andrew O'Keeffe, Draeger Medical Systems Inc
Geoff Pascoe
Joe Petruzzelli, Mindray DS USA Inc
Dewey Phan, Becton Dickinson & Company
Jon Platt, 3M Healthcare
Frank Pokrop, CareFusion
Bryan Pourciau, Cyberonics Inc
Inhel Rekik, University of Maryland Medical System
Victor Rodrigues, Cochlear Ltd
Albert Rodriguez, Cyberonics Inc
Miguel Rodriguez, Johnson & Johnson
Bill Roeca, CR Bard
Daniel Rubery, NxStage Medical Inc
Rick Schrenker, Massachusetts General Hospital
Ray Silkaitis, Amgen Inc

Chandresh Thakur, CareFusion
Donna-Bea Tillman, Biologics Consulting Group
Mark Walker, Focus Compliance & Validation Services
Janet Wickey, Bausch & Lomb Inc
Mike Willingham, Caradigm USA LLC
Daidi Zhong, Chongqing University

Alternates: Cheryl Carey, Medical Imaging & Technology Alliance a Division of NEMA
Kevin Crowe, Cochlear Ltd
Jean Desire, Hospira Worldwide Inc
Jeff Gilham, Spacelabs Healthcare
Jim Hempel, Covidien
Joshua Kim, Welch Allyn Inc
Don Lin
Don McAndrews, Philips Electronics North America
Andrew Northup, Medical Imaging & Technology Alliance a Division of NEMA
Susumu Nozawa, Becton Dickinson & Company
John Schmidt, Boston Scientific Corporation
Xianyu Shea, Stryker Medical Division
Diane Sheffer, Steris Corporation
Robert Smith, St Jude Medical Inc
Ferry Tamtoro, Amgen Inc
Ann Vu, Johnson & Johnson
Fei Wang, Fresenius Medical Care
J.S. Wiley, Draeger Medical Systems Inc
Xianyu Shea, Stryker Medical Division

NOTE—Participation by federal agency representatives in the development of this recommended practice does not constitute endorsement by the federal government or any of its agencies.

Foreword

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

DISCLAIMER

This Consolidated version is not an official IEC Standard and has been prepared for user convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.

This Consolidated version of IEC 62304 bears the edition number 1.1. It consists of the first edition (2006-05) [documents 62A/523/FDIS and 62A/528/RVD] and its amendment 1 (2015-06) [documents 62A/1007/FDIS and 62A/1014/RVD]. The technical content is identical to the base edition and its amendment.

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

International Standard IEC 62304 has been prepared by a joint working group of subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice and ISO Technical Committee 210, Quality management and corresponding general aspects for MEDICAL DEVICES. Table C.5 was prepared by ISO/IEC JTC 1/SC 7, Software and system engineering.

It is published as a dual logo standard.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard the following print types are used:

- requirements and definitions: in roman type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- terms used throughout this standard that have been defined in Clause 3 and also given in the index: in small capitals.

An asterisk (*) as the first character of a title or at the beginning of a paragraph indicates that there is guidance related to that item in Annex B.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment MANUFACTURERS and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.