

# Technical Information Report

## AAMI TIR17: 2017

Compatibility of materials  
subject to sterilization



# Compatibility of materials subject to sterilization

Approved 11 June 2017 by  
**Association for the Advancement of Medical Instrumentation**

**Abstract:** This technical information report provides guidance for health care product manufacturers in the qualification of polymeric materials, ceramics, and metals for use in health care products that are sterilized by the following modalities: a) radiation (gamma, electron beam, or x-ray); b) ethylene oxide; c) moist heat (steam); d) dry heat; e) hydrogen peroxide; f) nitrogen dioxide, g) peracetic acid vapor, h) liquid peracetic acid, and i) hydrogen peroxide–ozone. Annexes address the specific sterilization modality concerns.

**Keywords:** material qualification, sterilization

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

*Published by*

AAMI  
4301 N. Fairfax Dr., Ste. 301  
Arlington, VA 22203-1633

© 2018 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 4301 N. Fairfax Dr., Ste. 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 978-1-57020-700-6

This is a preview. [Click here to purchase the full publication.](#)

## Contents

Page

Committee representation.....	iv
Foreword.....	vii
1 Scope .....	1
2 Definitions, symbols, and abbreviations .....	2
3 Selection of materials.....	4
4 Manufacturing process and design considerations .....	33
5 Material testing.....	36
6 Accelerated aging programs.....	42
Annex A (informative) Radiation sterilization—Material compatibility fundamentals .....	43
Annex B (informative) Ethylene oxide sterilization—Material compatibility fundamentals .....	49
Annex C (informative) Moist heat sterilization—Material compatibility fundamentals.....	55
Annex D (informative) Dry heat sterilization—Material compatibility fundamentals .....	64
Annex E (informative) Hydrogen peroxide sterilization—Material qualification fundamentals .....	72
Annex F (informative) Nitrogen dioxide sterilization—Material qualification fundamentals .....	77
Annex G (informative) Peracetic acid (PA) vapor sterilization—Material compatibility fundamentals .....	82
Annex H (informative) Liquid peracetic acid sterilization—Material compatibility fundamentals.....	87
Annex I (informative) Hydrogen peroxide–ozone sterilization—Material compatibility fundamentals .....	91
Annex J (informative) Accelerated aging programs .....	95
Annex K (informative) Example of a device evaluation process .....	101
Annex L (informative) Material abbreviations .....	103
Bibliography .....	104

## Committee representation

### Association for the Advancement of Medical Instrumentation Compatibility of Materials Subject to Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Compatibility of Materials Subject to Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the TIR does not necessarily mean that all working group members voted for its approval. At the time this standard was published, the **AAMI Compatibility of Materials Subject to Sterilization Working Group** had the following members:

*Chair:* Karl Hemmerich

*Members:* Agnieszka Baczek, Medline Industries Inc  
Jenny Berg, Sterilucient Inc  
Carolyn Braithwaite-Nelson, Spectranetics Corporation  
Eunhee Cho, PhD, St Jude Medical Inc  
Nancy Chobin, RN, CSPM, CFER, Sterile Processing University LLC  
Sean Colwell, WuXi AppTec Inc  
Emily Craven, Mevex Corporation  
Chris Deneux, Becton Dickinson & Company  
John DiCaro, Medtronic Inc  
Mary Ann Drosnock, MS, Healthmark Industries Company Inc  
Gordon Ely, MiMedx Group  
Randy Eveland, PhD, STERIS Corporation  
Gloria Frost, PhD, Cardinal Health  
Joel Gorski, PhD, NAMSA  
Doug Harbrecht, Sterility Assurance LLC  
Arthur Harris, Cook Inc  
Fatima Hasanain, Sterigenics International  
Karl Hemmerich, Sterilization Validation Services  
Mollie Holter, Smiths Medical  
Nichole Jackson, Ecolab  
Nupur Jain, Intuitive Surgical Inc  
Carolyn Kinsley, LexaMed Ltd  
Ryan Klebba, Integrated Medical Systems  
Stacy Kromenhoek, Boston Scientific Corporation  
Byron Lambert, Abbott Laboratories  
Jean-Luc Lemyre, TSO<sub>3</sub> Inc  
Anne Lucas, PhD, FDA/CDRH  
Tania Lupu, Case Medical Inc  
Jeff Martin, Sterilization and Quality System Consulting LLC  
Gerry McDonnell, PhD, Johnson & Johnson  
Jami McLaren, PhD, Boston Scientific Corporation  
John Nedick, Olympus America Inc  
Gerry O'Dell, Gerry O'Dell Consulting  
Wayne Rogers, Wayne J Rogers Enterprises  
Mason Schwartz, Cantel Inc  
Paul Somodi, Hospira, a Pfizer company  
Larry Talapa, 3M Healthcare  
Don Tumminelli, HIGHPOWER Validation Testing & Lab Services Inc  
Wendy Wangsgard, Nelson Laboratories LLC  
Roberto Zumbado, Philips Electronics North America

*Alternates:* Jerome Bell, LexaMed Ltd  
Tim Carlson, Becton Dickinson & Company  
Peter Cheung, FDA/CDRH  
Alexandra Cooper, Arthrex Inc  
Mike DiCicco, PhD, Johnson & Johnson  
Dave Dion, Cardinal Health  
Chris Evans, Integrated Medical Systems  
Veronica Falkevitz, HIGHPOWER Validation Testing & Lab Services Inc

Niki Fidopiastis, Sterigenics International  
Elyse Gaudreau, TSO<sub>3</sub> Inc  
Betty Howard, STERIS Corporation  
Sandra Iverson, Cantel Inc  
Britt Jones, WuXi AppTec Inc  
Peter Kalkbrenner, Sterilucent Inc  
Kaumudi Kulkarni, Healthmark Industries Corporation  
Vu Le, Abbott Laboratories  
Nicole McLees, 3M Healthcare  
Michael O'Hara, FDA/CDRH  
Dave Parente, Ecolab  
Nicole Pasquino, Case Medical Inc  
Deanna Porter, St Jude Medical Inc  
Gary Socola, HIGHPOWER Validation Testing & Lab Services Inc  
Laxmishita Sreedasyam, Boston Scientific Corporation  
Mara Tafoya, WuXi AppTec Inc  
Brian Wallace, Intuitive Surgical Inc  
Martell Winters, Nelson Laboratories LLC

---

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

*Cochairs:* Michael H. Scholla, MS, PhD  
Patrick Weixel

*Members:* Anas Aljabo, PhD, SteriPro Canada Inc  
Brett Anderson, Cochlear Ltd  
Hank Balch, University Health System  
Richard Bancroft, STERIS Corporation  
Trabue D. Bryans, BryKor LLC  
Tim Carlson, Becton Dickinson & Company  
Phil Cogdill, Medtronic Inc  
Sean Colwell, WuXi AppTec Inc  
Ramona Conner, RN, MSN, CNOR, FAAN, Association of periOperative Registered Nurses  
Lena Cordie, Qualitas Professional Services LLC  
Jacqueline Daley, Sharp Metropolitan Medical Campus  
Gordon Ely, MiMedx Group  
Lisa Foster, Adiuvo QS & SA Consulting  
Joel R. Gorski, PhD, NAMSA  
Joyce Hansen, Johnson & Johnson  
Stephanie Homuth (Independent Expert)  
Clark Houghtling, Cosmed Group Inc  
Susan G. Klacik, CCSMC, FCS, ACE, International Association of Healthcare Central Service  
Materiel Management  
Byron J. Lambert, PhD, Abbott Laboratories  
Michelle Luebke, Baxter Healthcare Corporation  
Patrick J. McCormick, PhD, Bausch & Lomb Inc.  
Gerry McDonnell, PhD, Johnson & Johnson  
Gerry O'Dell, Gerry O'Dell Consulting  
Adrian Ponce, Verrix LLC  
Janet M. Prust, 3M Health Care  
Nancy J. Rakiewicz, IUVO BioScience  
Michael H. Scholla, MS, PhD, DuPont Protection Solutions  
Joan Spear, B Braun of America Inc  
Patrick Weixel, FDA/CDRH  
Sid Wiggs (Independent Expert)  
Martell Kress Winters, SM, Nelson Laboratories LLC  
Stephen Yeadon, Boston Scientific Corporation  
William E. Young, Sterigenics International  
Roberto Zumbado, Philips

*Alternates:* Stacy Bohl, Boston Scientific Corporation  
Jonathan Bull, Johnson & Johnson  
Greg Crego, IUVO BioScience  
Niki Fidopiastis, Sterigenics International  
Jeffrey Marx, STERIS Corporation  
Kimberly Patton, Becton Dickinson & Company  
Christine Render, Cosmed Group Inc  
Michael Sadowski, Baxter Healthcare Corporation  
Sharon Van Wicklin, Association of periOperative Registered Nurses  
Craig A. Wallace, 3M Health Care

---

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---



## Foreword

This AAMI technical information report (TIR) was developed to provide additional guidance in order to improve quality and reduce the costs and time required for performing material qualifications.

One of the activities encompassed within sterilization standards is to evaluate how the mode of sterilization affects product and packaging. This element is mentioned in each of the respective industrial sterilization standards (ANSI/AAMI/ISO 11135 series, ANSI/AAMI/ISO 11137 series, ANSI/AAMI/ISO 17665-1, and ANSI/AAMI/ISO 14937). The basic requirements of these standards include the implementation of a program to demonstrate the quality, safety, and performance of the product throughout its shelf life or until its expiration date. Components of such a program are 1) expeditious material selection, 2) prudent processing of those materials, 3) testing of any specific properties essential to the product's intended function, and 4) accelerated aging programs. AAMI TIR17:1997 addressed these four components of a material qualification program for radiation sterilization, and AAMI TIR17:2008 addressed these four components for additional sterilization modalities. There have been many requests from the health care manufacturing industry to expand material compatibility information to cover more sterilization modalities. Therefore, this TIR supersedes AAMI TIR17:2008, with an expanded scope that includes the following sterilization modalities:

- Radiation
- Ethylene oxide
- Moist heat (i.e., steam)
- Dry heat
- Hydrogen peroxide
- Nitrogen dioxide
- Peracetic acid vapor
- Liquid peracetic acid
- Hydrogen peroxide–ozone

These modalities are individually addressed in Section 3 and Annexes A through I of this TIR. Guidance on the processing of materials is carried over from AAMI TIR17:2008 and is provided in Section 4. General guidance on the testing of materials is provided in Section 5. Accelerated aging program information is provided in Section 6. If it has been carried over from AAMI TIR17:2008, or if it has been subsequently published elsewhere, references have been provided. To facilitate aging programs with the advent of combination devices, the accelerated aging information is supplemented with a comparison of accelerated aging programs for devices and accelerated stability programs for pharmaceuticals.

The bulk of the guidance on the compatibility of materials subject to sterilization is provided in Section 3 and the tables found in Annexes A through I. Each sterilization modality is described in enough detail (Section 3) for the reader to understand the parameters of the sterilization process that must be considered in evaluating material compatibility. Brief reference to the application of each sterilization modality to pharmaceutical and biological agents is also provided. One of the most beneficial aspects of the guidance in each Annex is a list of compatible materials to aid in the material selection process.

This TIR contains guidelines that are not intended to be absolute or applicable in all circumstances. Judgment should be used in applying the information in this TIR.

NOTE—This document is not an AAMI standard or an American National Standard, and the material contained herein is not normative in nature.

Suggestions for improving this TIR are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633.

---

NOTE—This foreword does not contain provisions of the AAMI technical information report, *Compatibility of materials subject to sterilization* (AAMI TIR17:2017), but it does provide important information about the development and intended use of the document.

---



# Compatibility of materials subject to sterilization

## 1 Scope

This document provides guidance for health care product manufacturers in the selection and qualification of polymeric materials, ceramics, and metals for use in health care products sterilized by the following methods:

- Radiation (gamma, electron beam, or x-ray)
- Ethylene oxide (EO)
- Moist heat (steam)
- Dry heat
- Hydrogen peroxide
- Nitrogen dioxide
- Vaporized peracetic acid
- Liquid peracetic acid
- Hydrogen peroxide–ozone

NOTE—All references to hydrogen peroxide sterilization in this TIR refer to sterilization in the gas phase. (Hydrogen peroxide is also used for liquid chemical sterilization, but that application is outside the scope of this TIR.)

Guidance in this TIR relates to the following:

- a) *Material selection*: Choosing sterilization-compatible materials (see Section 3 and Annexes A–I)
- b) *Material processing*: Optimizing the functional performance of materials selected to avoid processing errors that can contribute to negative effects from sterilization (see Section 4)
- c) *Material testing*: Challenging critical aspects of the product for functionality and safety after sterilization and aging (see Section 5)
- d) *Accelerated aging*: Applying programs that ensure correlation with real-time aging while reducing the cost and time required for material qualifications (see Section 6)

NOTE—The information in this TIR is not intended to provide a rationale for the use of materials without proper material qualification. The information is general and is intended only as a guide for successfully initiating material qualification programs.