

American National Standard

ANSI/AAMI ST91:2015

Flexible and semi-rigid
endoscope processing in
health care facilities

Objectives and uses of AAMI standards and recommended practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards healthcare professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the AAMI Vice President, Standards Policy and Programs. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

Flexible and semi-rigid endoscope processing in health care facilities

Approved 30 March 2015 by
Association for the Advancement of Medical Instrumentation

Approved 8 April 2015 by
American National Standards Institute Inc.

Abstract: Provides guidelines for precleaning, leak-testing, cleaning, packaging (where indicated), storage, high-level disinfecting, and/or sterilizing of flexible gastrointestinal (GI) endoscopes, flexible bronchoscopes, surgical flexible endoscopes (e.g., flexible ureteroscopes), and semi-rigid operative endoscopes (e.g., choledochoscopes) in health care facilities. These guidelines are intended to provide comprehensive information and direction for health care personnel in the processing of these devices and accessories.

Keywords: flexible endoscopes, high-level disinfection, semi-rigid endoscopes, sterilization

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Dr., Ste. 301
Arlington, VA 22203-1633

© 2015 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 4301 N. Fairfax Dr., Suite 301, Arlington, VA 22203-1633. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-585-X

This is a preview. [Click here to purchase the full publication.](#)

Contents

Page

| | |
|---|-----|
| Glossary of equivalent standards | vi |
| Committee representation | vii |
| Foreword | ix |
| Introduction | 1 |
| 1 Scope | 3 |
| 1.1 Inclusions | 3 |
| 1.2 Exclusions | 3 |
| 2 Definitions and abbreviations | 3 |
| 3 Design of endoscope processing area | 8 |
| 3.1 General considerations | 8 |
| 3.2 Work flow | 8 |
| 3.2.1 General considerations | 8 |
| 3.2.2 Physical separation | 8 |
| 3.2.3 Traffic control | 9 |
| 3.3 Physical facilities | 9 |
| 3.3.1 Space requirements | 9 |
| 3.3.2 Sinks and accessories | 10 |
| 3.3.3 Electrical systems | 10 |
| 3.3.4 Floors and walls | 10 |
| 3.3.5 Ceilings | 11 |
| 3.3.6 Doors | 11 |
| 3.3.7 Temperature, relative humidity, and ventilation | 11 |
| 3.3.8 Lighting | 12 |
| 3.3.9 Hand hygiene facilities | 13 |
| 3.3.10 Emergency eyewash/shower equipment | 13 |
| 3.3.11 Environmental cleaning | 14 |
| 4 Personnel | 14 |
| 4.1 General considerations | 14 |
| 4.2 Policies and procedures | 14 |
| 4.3 Education, training, and competency verification | 15 |
| 4.4 Standard precautions | 16 |
| 4.5 Hand hygiene | 17 |
| 4.6 Attire | 17 |
| 4.6.1 General considerations | 17 |
| 4.6.2 Personal protective equipment | 17 |
| 4.7 Immunizations | 18 |
| 5 Cleaning and high-level disinfection | 19 |
| 5.1 General considerations | 19 |
| 5.2 Precleaning at the point of use | 19 |
| 5.3 Transporting used endoscopes | 19 |
| 5.4 Leak testing | 20 |
| 5.4.1 General considerations | 20 |
| 5.4.2 Manual (dry) leak testing | 20 |

| | | |
|--------|--|----|
| 5.4.3 | Mechanical (wet) leak testing | 20 |
| 5.4.4 | Mechanical (dry) leak testing | 21 |
| 5.4.5 | Mechanical leak testing using an AER | 21 |
| 5.4.6 | Leak test failures | 22 |
| 5.5 | Manual cleaning..... | 22 |
| 5.6 | Manual rinsing | 23 |
| 5.7 | High-level disinfection and liquid chemical sterilization..... | 23 |
| 5.7.1 | General considerations..... | 23 |
| 5.7.2 | High-level disinfection..... | 23 |
| 5.7.3 | Liquid chemical sterilants/liquid chemical sterilization..... | 24 |
| 5.7.4 | Manual liquid chemical sterilization/high-level disinfection | 25 |
| 6 | Automated endoscope reprocessors | 26 |
| 7 | Sterile endoscope sheaths used as protective microbial barriers | 28 |
| 8 | Terminal sterilization by gaseous chemical sterilization processes..... | 28 |
| 8.1 | General considerations | 28 |
| 8.2 | Packaging for terminal sterilization | 29 |
| 8.2.1 | General considerations..... | 29 |
| 8.2.2 | Sterilization pouches | 29 |
| 8.2.3 | Sterilization wraps | 29 |
| 8.2.4 | Rigid sterilization containment systems..... | 29 |
| 8.3 | Ethylene oxide gas (EO) sterilization | 29 |
| 8.4 | Hydrogen peroxide gas sterilization..... | 30 |
| 8.5 | Ozone sterilization | 31 |
| 9 | Processing of endoscope accessories..... | 31 |
| 10 | Storage of reprocessed endoscopes | 32 |
| 10.1 | General Considerations | 32 |
| 10.2 | Storage of high-level disinfected endoscopes..... | 33 |
| 10.3 | Storage of sterilized endoscopes..... | 33 |
| 10.4 | Hang time for high-level disinfected endoscopes..... | 33 |
| 10.4.1 | General considerations..... | 33 |
| 10.4.2 | Existing guidelines..... | 33 |
| 10.4.3 | Risk assessment | 34 |
| 11 | Transport of high-level disinfected endoscopes | 35 |
| 12 | Quality Control | 35 |
| 12.1 | General considerations | 35 |
| 12.2 | Product identification and traceability..... | 37 |
| 12.3 | Documentation and record-keeping | 37 |
| 12.3.1 | Documentation | 37 |
| 12.3.2 | Expiration dating..... | 38 |
| 12.4 | Verification and monitoring of the cleaning process..... | 38 |
| 12.4.1 | General Considerations | 38 |
| 12.4.2 | Cleaning verification..... | 39 |
| 12.4.3 | Cleaning verification tests for users | 39 |
| 12.4.4 | Testing cleaning efficacy | 40 |
| 12.5 | Monitoring processes that use liquid chemical sterilization/high-level disinfection..... | 40 |
| 12.5.1 | Manual processes | 40 |

| | | |
|---------|--|----|
| 12.5.2 | Automated processes..... | 42 |
| 12.6 | Monitoring gaseous chemical sterilization processes | 44 |
| 12.6.1 | Use of physical monitors | 44 |
| 12.6.2 | Gaseous chemical sterilizer malfunction | 44 |
| 12.7 | Chemical indicators..... | 45 |
| 12.7.1 | General considerations..... | 45 |
| 12.7.2 | Using chemical indicators..... | 45 |
| 12.7.3 | Nonresponsive or inconclusive chemical indicators..... | 45 |
| 12.8 | Biological indicators | 46 |
| 12.8.1 | General considerations..... | 46 |
| 12.8.2 | Using biological indicators and process challenge devices | 46 |
| 12.8.3 | Frequency of use of BIs and PCDs | 46 |
| 12.9 | Sterilizer testing | 46 |
| 12.9.1 | General considerations..... | 46 |
| 12.9.2 | Qualification test procedure with BIs | 47 |
| 12.9.3 | Qualification testing acceptance criteria | 47 |
| 12.9.4 | Routine test procedure with BIs..... | 47 |
| 12.9.5 | Routine testing acceptance criteria..... | 47 |
| 12.9.6 | Positive BI results | 47 |
| 12.9.7 | Microbiological testing | 48 |
| 12.9.8 | Product release | 48 |
| 12.10 | Product recalls | 48 |
| 12.10.1 | General considerations..... | 48 |
| 12.10.2 | Recall procedure | 49 |
| 12.10.3 | Recall order | 49 |
| 12.10.4 | Recall summary report | 49 |
| 12.10.5 | Outbreak report | 49 |
| 12.11 | Quality process improvement | 49 |
| 12.11.1 | General considerations..... | 49 |
| 12.11.2 | Risk analysis..... | 50 |
| 12.11.3 | Decontamination..... | 51 |
| 12.11.4 | Liquid chemical sterilization, high-level disinfection, and gaseous chemical sterilization | 51 |
| 12.11.5 | Functional areas for product and process improvement..... | 52 |
| | Bibliography | 54 |

Tables

| | | |
|---|--|----|
| 1 | IES-recommended illuminance levels for work environments | 22 |
| 2 | Endoscope storage risk assessment checklist..... | 44 |

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf