



AGA Report No. 9

Measurement of Gas by Multipath Ultrasonic Meters

**Second Edition
April 2007**



AGA Report No. 9

Measurement of Gas by Multipath Ultrasonic Meters

Second Edition

April 2007

Transmission Measurement Committee



Copyright 2007, Operating Section, American Gas Association
400 North Capitol Street, NW, 4th Floor, Washington, DC 20001, U.S.A.
Phone: (202) 824-7000 • Fax: (202) 824-7082 • Web: www.aga.org

Catalog # XQ0701

DISCLAIMERS AND COPYRIGHT

The American Gas Association's (AGA) Operating Section provides a forum for industry experts to bring collective knowledge together to improve the state of the art in the areas of operating, engineering and technological aspects of producing, gathering, transporting, storing, distributing, measuring and utilizing natural gas.

Through its publications, of which this is one, AGA provides for the exchange of information within the gas industry and scientific, trade and governmental organizations. Each publication is prepared or sponsored by an AGA Operating Section technical committee. While AGA may administer the process, neither AGA nor the technical committee independently tests, evaluates or verifies the accuracy of any information or the soundness of any judgments contained therein.

AGA disclaims liability for any personal injury, property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of or reliance on AGA publications. AGA makes no guaranty or warranty as to the accuracy and completeness of any information published therein. The information contained therein is provided on an "as is" basis and AGA makes no representations or warranties including any expressed or implied warranty of merchantability or fitness for a particular purpose.

In issuing and making this document available, AGA is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is AGA undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

AGA has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does AGA list, certify, test or inspect products, designs or installations for compliance with this document. Any certification or other statement of compliance is solely the responsibility of the certifier or maker of the statement.

AGA does not take any position with respect to the validity of any patent rights asserted in connection with any items that are mentioned in or are the subject of AGA publications, and AGA disclaims liability for the infringement of any patent resulting from the use of or reliance on its publications. Users of these publications are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Users of this publication should consult applicable federal, state and local laws and regulations. AGA does not, through its publications intend to urge action that is not in compliance with applicable laws, and its publications may not be construed as doing so.

This report is the cumulative result of years of experience of many individuals and organizations acquainted with the measurement of natural gas. However, changes to this report may become necessary from time to time. If changes to this report are believed appropriate by any manufacturer, individual or organization, such suggested changes should be communicated to AGA by completing the last page of this report titled, **"Form for Proposal on AGA Report No. 9"** and sending it to: **Operations & Engineering Services Group, American Gas Association, 400 North Capitol Street, NW, 4th Floor, Washington, DC 20001, U.S.A.**

Copyright 2007, American Gas Association, All Rights Reserved.

FOREWORD

This report is published in the form of a performance-based specification for multipath ultrasonic meters for gas flow measurement. It is the result of a collaborative effort of users, meter manufacturers, flow measurement research organizations and independent consultants forming Task Group R-9 of AGA's Transmission Measurement Committee (TMC). In addition, comments to this report were made by the Committee on Gas Flow Measurement (COGFM) of the American Petroleum Institute (API), Gas Processor's Association (GPA), International Standard Organization's ISO/TC 30/SC 5/WG 1, Pipeline Research Council International (PRCI) and the committee members of International School of Hydrocarbon Measurement (ISHM).

This version of AGA Report No. 9 is intended to supersede all prior versions of this document. However, this document does not reference existing multipath ultrasonic meter installations. The decision to apply this document to existing installations shall be at the discretion of the parties involved.

Research conducted in support of this report and cited herein has demonstrated that multipath ultrasonic meters can accurately measure gas flow and, therefore, should be able to meet or exceed the requirements specified in this report when calibrated and installed according to the recommendations contained herein. Users should follow appropriate installation, use and maintenance of an ultrasonic meter as applicable in each case.

Various combinations of upstream fittings, valves and lengths of straight pipe can produce profile disturbances at the meter inlet that may result in flow-rate measurement errors. The amount of meter error will depend on the magnitude of the inlet velocity profile distortion produced by the upstream piping configuration and the meter's ability to compensate for this distortion. Research results and flow-meter calibration data have indicated that multipath ultrasonic flow meters can accurately measure gas flow rate when installed with upstream piping lengths and/or flow conditioning systems sufficient to maintain the integrity of the flow calibration. Other effects that may also result in flow-rate measurement errors for a given installation include levels of pulsation, range of operating pressures and ambient temperature conditions.

Flow-calibration guidelines are provided for occasions when a flow calibration is requested or required to verify the meter's accuracy or to apply a calibration factor to minimize the measurement uncertainty. (See Report text and Appendix A)

Unlike most traditional gas meters, multipath ultrasonic meters inherently have an embedded microprocessor system. Therefore, this report includes, by reference, a standardized set of international testing specifications applicable to electronic gas meters. These tests, summarized in Appendix B, are used to demonstrate the acceptable performance of the multipath ultrasonic meter's electronic system design under different influences and disturbances.

AGA Engineering Technical Note M-96-2-3, Ultrasonic Flow Measurement for Natural Gas Applications, is included in Appendix C, as a source of background information on ultrasonic gas metering. Contents of this technical note were based on the information available when the note was written in March 1996. Therefore, in case of any conflict between the information in the main report and the technical note (Appendix C), the content in the main report prevails.

The flow meter and/or flow conditioner performance verification test found in Appendix D is intended to provide a method by which an ultrasonic flow metering system can be shown to perform acceptably (i.e., within the performance specifications described in Sections 5.1, 5.1.1 and 5.1.2 of this document) under varying test flow conditions.

An example of overall measurement uncertainty calculations is provided in Appendix E with assumed numerical values for estimating measurement uncertainty for sites using ultrasonic gas flow meters.

ACKNOWLEDGMENTS

AGA Report No. 9, *Measurement of Gas by Multipath Ultrasonic Meters*, was revised by a Task Group of the American Gas Association's Transmission Measurement Committee under the joint chairmanship of **Paul LaNasa** of CPL & Associates and **Warren Peterson** of TransCanada Corporation. Individuals who made substantial contributions to the revision of this document are:

Jim Bowen, formerly of Instromet, Inc.
Ed Bowles, Southwest Research Institute
Joe Bronner, Pacific Gas & Electric Co
Larry Fraser, Fraser & Associates
Garnet Grudeski, TransCanada Calibrations
John Lansing, Sick - Maihak
Dan Rebman, Columbia Gulf Transmission Co.
Jim Witte, El Paso Pipeline Group

Other individuals who contributed to the revision of the document are:

Frank Brown, Consultant
Cary Carter, Texas Gas Transmission
Claire Becker-Castle, Sempra Utilities
Craig Chester, Williams Gas Pipeline
Joel Clancy, CEESI
Peter Espina, Controlotron Corporation
Angela Floyd, Panhandle Energy
Bill Frasier, Northern Border Pipeline
Robert Fritz, Lone Star Measurement
Jim Griffeth, Bristol Babcock, Inc.
Terrence Grimley, Southwest Research Institute
Danny Harris, Columbia Gas
Zaki Husain, Chevron Texaco
Mark Imboden, Controlotron Corporation
Jim Keating, Consultant
Allen Knack, Consumers Energy
Rick Ledesma, El Paso Pipeline Group
Brad Massey, Southern Star Central Gas Pipeline
George Mattingly, Consultant
Dannie Mercer, Atmos Energy
Kevin Moir, DTE Energy
Dr. Thomas Morrow, Southwest Research Institute
John Naber, Emerson Process Management, Daniel Div.
Dan Peace, Sensus Metering Systems
Mark Pelkey, National Fuel Gas
Thanh Phan, Duke Energy
Reese Platzer, Questar Pipeline
Alex Podgers, American Meter Co.
Hank Poellnitz, III, El Paso Pipeline Group
King Poon, Thermo Electron Corporation
James Robertson, Pacific Gas & Electric Co.
Blaine Sawchuk, Canada Pipeline Accessories
Mike Scelzo, GE Sensing

Walt Seidl, CEESI
Tushar Shah, Eagle Research Corporation
Jerry Paul Smith, Consultant
Karl Stappert, Daniel M&C
John Stuart, Stuart Gas Measurement Consulting
Fred VanOrsdol, Southern Petroleum Laboratories
Kevin Warner, Instromet, Inc.
Wayne Wenger, Kinder Morgan
Klaus Zanker, Emerson Process Management, Daniel Div.

AGA acknowledges the contributions of the above individuals and thanks them for their time and effort in getting this document revised.

Lori Traweck
Senior Vice President

Ali Quraishi
Staff Executive, Engineering Services Director

Table of Contents

1.	INTRODUCTION	1
1.1	Scope.....	1
1.2	Principle of Measurement.....	1
2.	TERMINOLOGY, UNITS AND DEFINITIONS.....	2
2.1	Terminology	2
2.2	Engineering Units	2
2.3	Definitions	3
3.	OPERATING CONDITIONS.....	6
3.1	Gas Quality	6
3.2	Pressures	6
3.3	Temperatures, Gas and Ambient.....	6
3.4	Gas Flow Considerations	6
3.5	Upstream Piping and Flow Profiles	7
3.6	Acoustic Noise	7
4.	METER REQUIREMENTS	9
4.1	Codes and Regulations.....	9
4.2	Quality Assurance	9
4.3	Meter Body.....	9
4.3.1	Maximum Operating Pressure	9
4.3.2	Corrosion Resistance	9
4.3.3	Meter Body Lengths and Bores.....	9
4.3.4	Ultrasonic Transducer Ports	10
4.3.5	Pressure Tap	10
4.3.6	Miscellaneous	10
4.3.7	Meter Body Markings	11
4.4	Ultrasonic Transducers.....	11
4.4.1	Specifications	11
4.4.2	Rate of Pressure Change.....	11

4.4.3	Exchange	11
4.4.4	Transducer Tests	12
4.5	Electronics	12
4.5.1	General Requirements	12
4.5.2	Output Signal Specifications	12
4.5.3	Electrical Safety Design Requirements	13
4.5.4	Component Replacement	13
4.6	Computer Programs	13
4.6.1	Firmware	13
4.6.2	Configuration and Maintenance Software	14
4.6.3	Inspection and Auditing Functions	14
4.6.4	Alarms	14
4.6.5	Diagnostic Measurements	14
4.7	Documentation	15
4.7.1	After Receipt of Order	16
4.7.2	Before Shipment	16
5.	METER PERFORMANCE REQUIREMENTS	17
5.1	General Meter Performance Requirements	17
5.1.1	Large Meter Accuracy	17
5.1.2	Small Meter Accuracy	18
5.2	Pressure, Temperature and Gas Composition Influences	19
6.	INDIVIDUAL METER & METERING PACKAGE TESTING REQUIREMENTS....	20
6.1	Leakage Tests	20
6.2	Dimensional Measurements	20
6.3	Zero-Flow Verification Test (Zero Test)	21
6.4	Meter and Metering Package Flow-Calibration Test	21
6.4.1	Preparation for Flow Calibration	22
6.4.2	Calibration of Metering Package	23
6.4.3	Calibration Adjustment Factors	23
6.4.4	Calibration Test Reports	23
6.4.5	Final Considerations	24
7.	INSTALLATION REQUIREMENTS	25
7.1	Environmental Considerations	25
7.1.1	Temperature	25
7.1.2	Vibration	25
7.1.3	Electrical Noise	25
7.1.4	Pulsation	25