



Designation: E1193 – 20

Standard Guide for Conducting *Daphnia magna* Life-Cycle Toxicity Tests¹

This standard is issued under the fixed designation E1193; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ε) indicates an editorial change since the last revision or reappraisal.

1. Scope

1.1 This guide covers procedures for obtaining laboratory data concerning the adverse effects of a test material (added to dilution water, but not to food) on *Daphnia magna* Straus, 1820, during continuous exposure throughout a life-cycle using the renewal or flow-through techniques. These procedures also should be useful for conducting life-cycle toxicity tests with other invertebrate species and cladocerans from the same genus (for example, *Daphnia pulex*), although modifications might be necessary.

1.2 These procedures are applicable to most chemicals, either individually or in formulations, commercial products, or known mixtures. With appropriate modifications, these procedures can be used to conduct tests on temperature, dissolved oxygen, pH, and on such materials as aqueous effluents (also see Guide E1192), leachates, oils, particulate matter, sediments, and surface waters. The technique, (renewal or flow-through), will be selected based on the chemical characteristics of the test material such as high oxygen demand, volatility, susceptibility to transformation (biologically or chemically), or sorption to glass.

1.3 Modification of these procedures might be justified by special needs or circumstances. Although using appropriate procedures is more important than following prescribed procedures, results of tests conducted using unusual procedures are not likely to be comparable to results of standard test procedures. Comparison of results obtained using modified and unmodified versions of these procedures might provide useful information on new concepts and procedures for conducting life-cycle toxicity tests with *D. magna*. Appendix X3 provides modifications for conducting the chronic toxicity test method with *D. pulex* Leydig, 1860.

1.4 This guide is arranged as follows:

Section

Referenced Documents	2
Terminology	3
Summary of Guide	4
Significance and Use	5
Apparatus	6
Facilities	6.1
Construction Materials	6.2
Test Chambers	6.3
Cleaning	6.4
Acceptability	6.5
Reagents	7
Purity of Reagents	7.1
Hazards	8
Dilution Water	9
Requirements	9.1
Source	9.2
Treatment	9.3
Characterization	9.4
Test Material	10
General	10.1
Stock Solutions	10.2
Test Concentrations(s)	10.3
Test Organisms	11
Species	11.1
Age	11.2
Source	11.3
Brood Stock	11.4
Food	11.5
Handling	11.6
Harvesting Young	11.7
Quality	11.8
Procedure	12
Experimental Design	12.1
Dissolved Oxygen	12.2
Temperature	12.3
Loading	12.4
Selection of Test System	12.5
Beginning the Test	12.6
Care and Maintenance	12.7
Feeding	12.8
Duration	12.9
Biological Data	12.10
Other Measurements	12.11
Analytical Methodology	13
Acceptability of Test	14
Calculation of Results	15
Report	16
Keywords	17
Appendices	
Appendix X1 Statistical Guidance	
Appendix X2 Food	
Appendix X3 Modifications for Conducting Chronic Life Cycle Analysis Tests with <i>Daphnia Pulex</i>	

¹ This guide is under the jurisdiction of ASTM Committee E50 on Environmental Assessment, Risk Management and Corrective Action and is the direct responsibility of Subcommittee E50.47 on Biological Effects and Environmental Fate.

Current edition approved Dec. 1, 2020. Published January 2021. Originally approved in 1987. Last previous edition approved in 2012 as E1193 – 97 (2012). DOI: 10.1520/E1193-20.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* Specific hazard statements are given in Section 8.

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

[E729 Guide for Conducting Acute Toxicity Tests on Test Materials with Fishes, Macroinvertebrates, and Amphibians](#)

[E943 Terminology Relating to Biological Effects and Environmental Fate](#)

[E1023 Guide for Assessing the Hazard of a Material to Aquatic Organisms and Their Uses](#)

[E1192 Guide for Conducting Acute Toxicity Tests on Aqueous Ambient Samples and Effluents with Fishes, Macroinvertebrates, and Amphibians](#)

[E1733 Guide for Use of Lighting in Laboratory Testing](#)

[E1847 Practice for Statistical Analysis of Toxicity Tests Conducted Under ASTM Guidelines](#)

[IEEE/ASTM SI 10 American National Standard for Use of the International System of Units \(SI\): The Modern Metric System](#)

3. Terminology

3.1 The words “must,” “should,” “may,” “can,” and “might” have very specific meanings in this guide.

3.2 *must*—used to express an absolute requirement, that is, to state that the test ought to be designed to satisfy the specified condition, unless the purpose of the test requires a different design. “Must” is used only in connection with factors that directly relate to the acceptability of the test (see [14.1](#)).

3.3 *should*—used to state that the specified condition is recommended and ought to be met if possible. Although violation of one “should” is rarely a serious matter, violation of several will often render the results questionable. Terms such as “is desirable,” “is often desirable,” and “might be desirable” are used in connection with less important factors.

3.4 *may*—used to mean “is (are) allowed to,” “can” is used to mean “is (are) able to,” and “might” is used to mean “could possibly.” Therefore the classic distinction between “may” and

“can” is preserved, and “might” is never used as a synonym for either “may” or “can.”

3.5 For definitions of other terms used in this guide, refer to Guide [E729](#) and Terminology [E943](#). For an explanation of units and symbols, refer to [IEEE/ASTM SI 10](#).

4. Summary of Guide

4.1 A 21-day life-cycle toxicity test for *Daphnia magna* is described. The test design allows for the test organisms to be exposed to a test material using either the renewal technique (with exchange of the total volume of test water and toxicant at least three times a week) or the flow-through technique (with continual water and toxicant addition, usually at least four volume additions per day). At least five concentrations of a test material, a control, and a solvent control (if applicable) replicated at least four times are recommended. Each test concentration has at least ten *Daphnia* per treatment. There are two applicable exposure approaches (both may be renewal or flow-through): (1) use of a minimum of ten daphnids per treatment and one daphnid per replicate (renewal or flow-through); or (2) use of four replicates with at least five daphnids per replicate (≥ 20 daphnids per treatment). A control consists of maintaining daphnids in dilution water to which no test material has been added to provide (1) a measure of the acceptability of the test by giving an indication of the quality of the test organisms and the suitability of the dilution water, food, test conditions, handling procedures, and so forth, and (2) the basis for interpreting data obtained from the other treatments. In each of the other treatments, the daphnids are maintained in dilution water, to which a selected concentration of test material has been intentionally added. Measurement end points obtained during the test include the concentration of the test material and final number alive, final weight, final length, and number of progeny per daphnid. Data are analyzed to determine the effect of the test material on survival, growth, and reproduction of *D. magna*. These methods are also applicable for chronic toxicity testing using the conspecific *Daphnia pulex*, with slight modifications provided in [Appendix X3](#).

5. Significance and Use

5.1 Protection of an aquatic species requires prevention of unacceptable effects on populations in natural habitats. Toxicity tests are conducted to provide data that may be used to predict what changes in numbers and weights of individuals might result from similar exposure to the test material in the natural aquatic environment. Information might also be obtained on the effects of the material on the health of the species.

5.2 Results of life-cycle tests with *D. magna* are used to predict chronic effects likely to occur on daphnids in field situations as a result of exposure under comparable conditions.

5.2.1 Life-cycle tests with *D. magna* are used to compare the chronic sensitivities of different species, the chronic toxicities of different materials, and study the effects of various environmental factors on the results of such tests.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

5.2.2 Life-cycle tests with *D. magna* are used to assess the risk of materials to aquatic organisms (see Guide E1023) or derive water quality criteria for aquatic organisms (1).³

5.2.3 Life-cycle tests with *D. magna* are used to extrapolate the results of chronic toxicity tests on the same test material with the same species in another water or with another species in the same or a different water. Most such predictions take into account the results of acute toxicity tests, and so the usefulness of the results of a life-cycle test with *D. magna* may be increased by reporting the results of an acute toxicity test (see Guide E729) conducted under the same conditions. In addition to conducting an acute toxicity test with unfed *D. magna*, it may be relevant to conduct an acute test in which the daphnids are fed the same as in the life-cycle test to see if the presence of that concentration of that food affects the results of the acute test and the acute-chronic ratio (ACR) (see 10.3.1).

5.2.4 Life-cycle tests are used to evaluate the biological availability of, and structure-activity relationships between, test materials and test organisms.

5.3 Results of life-cycle tests with *D. magna* might be influenced by temperature (2), quality of food, composition of dilution water, condition of test organisms, test media (for example, water hardness), and other factors.

6. Apparatus

6.1 *Facilities*—Culture and test chambers are often kept in a separate room maintained at about 20 °C. Alternatively, cultures and test chambers may be placed in a temperature-controlled water bath, environmental chamber, or incubator. The water-supply system should provide an adequate supply of dilution water to the culture tanks and test chambers. The water-supply system should be equipped for temperature control and aeration, and strainers and air traps should be included in the water-supply system. Air used for aeration should be free of fumes, oil, and water; filters to remove oil and water are desirable. Filtration of air through a 0.22-µm bacterial filter might be desirable (3). During culturing and testing, daphnids should be shielded from disturbances to prevent unnecessary stress. The test facility should be well-ventilated and free of fumes. A timing device should be used to provide a 16-h light and 8-h dark photoperiod (4). A 15 to 30-min transition period when lights go on might be desirable to reduce the possibility of daphnids being stressed by instantaneous illumination; a transition period when lights go off may also be desirable (see Guide E1733).

6.1.1 When *D. magna* are fed live algae, a high-light intensity might cause sufficient photosynthesis to result in an increase of pH high enough to kill daphnids (5). Therefore, the maximum acceptable intensity is dependent on the buffer capacity of the dilution water, species, and density of algae, and the kind of test chamber and cover (see Guide E1733). Light intensities up to 600 lx or a fluence rate of 1 w/m² will usually be acceptable, but higher intensities might result in an unacceptably high pH in the culture water.

6.2 *Construction Materials*—Equipment and facilities that contact stock solutions, test solutions, or any water into which daphnids will be placed should not contain substances that can be leached or dissolved by aqueous solutions in amounts that can adversely affect daphnids. In addition, equipment and facilities that contact stock solutions or test solutions should be chosen to minimize sorption of test materials from water. Glass, Type 316 stainless steel, nylon, fiberglass, silicon, and fluorocarbon plastics should be used whenever possible to minimize leaching, dissolution, and sorption. Concrete and rigid (unplasticized) plastics may be used for culture tanks and in the water-supply system, but they should be soaked, preferably in flowing dilution water, for several days before use (6). Cast-iron pipe may be used in supply systems, but colloidal iron probably will be added to the dilution water and strainers will be needed to remove rust particles. Copper, brass, lead, galvanized metal, and natural rubber should not contact dilution water, stock solutions, or test solutions before or during the test. Items made of neoprene rubber and other materials not previously mentioned should not be used unless it has been shown that their use will not adversely affect survival, growth, and reproduction of *D. magna* (see Section 14).

6.3 Test Chambers:

6.3.1 *Flow-through tests*, 500-mL to 2-L glass beakers (or equivalent) with a notch (approximately 4 by 13 cm) cut in the lip may be used to expose the *Daphnia* to the test material. The notch should be covered with 0.33-mm opening (U.S. standard sieve size No. 50) stainless steel or polyethylene screening small enough to retain first instar *Daphnia*. The screen can be attached to the beaker with silicone adhesive. The chambers should provide at least 30 mL of solution for each of the initial test daphnid(s).

6.3.2 *Renewal tests*, beaker ranging in size from 100 to 1000 mL. A notched chamber is not required for a renewal test. Each chamber should provide at least 40 mL of solution for each of the initial test daphnid(s).

6.3.3 Any container made of glass, Type 316 stainless steel, or a fluorocarbon plastic may be used if (1) each chamber is separate with no interconnections, (2) each chamber contains at least 30 mL of test solution (see 12.4) per first-generation daphnid for flow-through tests and at least 40 mL for renewal tests, (3) there is at least 1000 mm² of air to water interface per daphnid, and (4) the test solution is at least 30 mm deep. Static test chambers should be covered with glass, stainless steel, nylon, or fluorocarbon plastic covers to keep out extraneous contaminants and to reduce evaporation of test solution. All chambers and covers in a test must be identical. Covers are not required for flow-through studies.

6.4 *Cleaning*—Test chambers and equipment used to prepare and store dilution water, stock solutions, and test solutions should be cleaned before use. New equipment should be washed with detergent and rinsed with water, a water-miscible organic solvent, water, acid (such as 5 % concentrated nitric acid), and washed at least twice with distilled, deionized, or dilution water. Some lots of some organic solvents might leave a film that is insoluble in water. Also, stronger nitric acid, for example, 10 %, might cause deterioration of silicone adhesive; an initial rinse with 10 % concentrated hydrochloric acid might

³ The boldface numbers in parentheses refer to the list of references at the end of this guide.

prevent such deterioration. A dichromate-sulfuric acid cleaning solution can generally be used in place of both the organic solvent and the acid, but it might attack silicone adhesives. At the end of every test, all items that are to be used again should be immediately (1) emptied, (2) rinsed with water, (3) cleaned by a procedure appropriate for removing the test material (for example, acid to remove metals and bases; detergent, organic solvent, or activated carbon to remove organic chemicals), and (4) rinsed at least twice with distilled, deionized, or dilution water. Acid is useful for removing mineral deposits. Test chambers should be rinsed with dilution water just before use.

6.5 Acceptability—Before a toxicity test is conducted in new test facilities, it is desirable to conduct a control water-only (that is, no test material) test, in which all test chambers contain dilution water. This test will reveal (1) whether *D. magna* will survive, grow, and reproduce acceptably (see Section 14) in the new facilities, (2) whether there are any location effects on survival, growth, or reproduction, and (3) the magnitude of the within-chamber and between-chamber variance.

7. Reagents

7.1 Purity of Reagents—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the test.

8. Hazards

8.1 Many materials can affect humans adversely if precautions are inadequate. Therefore, skin contact with all test materials and solutions should be minimized by wearing appropriate protective gloves (especially when washing equipment or putting hands in test solutions), laboratory coats, aprons, and glasses, and by using dip nets or tubes to remove daphnids from test solutions. Special precautions, such as covering test chambers and ventilating the area surrounding the chambers, should be taken when conducting tests on volatile materials. Information on toxicity to humans (7), recommended handling procedures (8), and chemical and physical properties of the test material should be studied before a test is begun. Special procedures will be necessary with radiolabeled test materials (9) and with materials that are, or are suspected of being, carcinogenic (10).

8.2 Disposal of stock solutions, test solutions, and test organisms might pose special problems in some cases; therefore, health and safety precautions and applicable regulations should be considered before beginning a test. Removal or

degradation of test material might be desirable before disposal of stock and test solutions.

8.3 Cleaning equipment with a volatile solvent such as acetone should be performed only in a well-ventilated area with no smoking allowed and no open flame, for example, pilot light, is present.

8.4 Acidic solutions and hypochlorite solutions should not be mixed together because hazardous fumes might be produced.

8.5 Because dilution water and test solutions are usually good conductors of electricity, use of ground fault systems and leak detectors should be considered to help prevent electrical shocks.

8.6 To prepare dilute acid solutions, concentrated acid should be added to water, not vice versa. Opening a bottle of concentrated acid and mixing concentrated acid with water should be performed only in a well-ventilated area.

9. Dilution Water

9.1 Requirements—The dilution water should (1) be acceptable to *D. magna*, (2) be of uniform quality, and (3), except as stated in 9.1.4, not unnecessarily affect results of the test.

9.1.1 The dilution water must allow satisfactory survival, growth, and reproduction of *D. magna* (see Section 14).

9.1.2 The quality of the dilution water should be uniform, allowing the brood stock to be cultured and the test conducted in water of the same quality. In particular, during culture or testing, or both, the range of hardness should be $\pm 10\%$ of the average.

9.1.3 The dilution water should not unnecessarily affect results of a life-cycle test with *D. magna* because of such things as sorption or complexation of test material. Therefore, except as stated in 9.1.4, concentrations of both total organic carbon (TOC) and particulate matter should be less than 5 mg/L.

9.1.4 If it is desired to study the effect of an environmental factor such as TOC, particulate matter, or dissolved oxygen on the results of a life-cycle test with *D. magna*, it will be necessary to use a water that is naturally or artificially high in TOC or particulate matter or low in dissolved oxygen. If such a water is used, it is important that adequate analyses be performed to characterize the water, and that a comparable test be available or conducted in the laboratory's usual culture dilution water to facilitate interpretation of the results in the special water.

9.2 Source:

9.2.1 The use of reconstituted water might increase comparability of test results between laboratories. The hard reconstituted fresh water (160 to 180 mg/L as CaCO_3) described in Guide E729 has been used successfully. Addition of 2 μg of selenium(IV) and 1 μg of crystalline vitamin B_{12} /L might be desirable (11). Other water sources (natural or reconstituted) may be used if they have been demonstrated to provide adequate daphnid survival, growth, and reproduction.

9.2.2 Natural fresh waters have been used successfully. Natural waters should be obtained from an uncontaminated source of consistent quality. A well or spring is usually

⁴ Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.