



BSI Standards Publication

In vitro diagnostic medical devices — Clinical performance studies using specimens from human subjects — Good study practice

National foreword

This British Standard is the UK implementation of [ISO 20916:2019](#).

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2019
Published by BSI Standards Limited 2019

ISBN 978 0 580 97940 8

ICS 11.100.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2019.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

INTERNATIONAL STANDARD

ISO
20916

First edition
2019-05-31

In vitro diagnostic medical devices — Clinical performance studies using specimens from human subjects — Good study practice

*Dispositifs médicaux de diagnostic in vitro — Études des
performances cliniques utilisant des prélèvements de sujets humains
— Bonnes pratiques d'étude*



Reference number
ISO 20916:2019(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Ethical considerations	11
4.1 General	11
4.2 Improper influence or inducement	11
4.3 Responsibilities	11
4.4 Ethics committee involvement	11
4.5 Informed consent	12
5 Clinical performance study planning	12
5.1 General	12
5.2 Risk evaluation	13
5.3 Design of the clinical performance study	14
5.4 Investigator brochure	14
5.5 Clinical Performance Study Protocol (CPSP)	15
5.5.1 General	15
5.5.2 Principal investigator responsibilities	15
5.5.3 Contents of the CPSP	16
5.6 Case report forms	19
5.7 Recording of specimen information	20
5.8 Specimen accountability and integrity	20
5.9 Study site selection	20
5.9.1 Site qualification	20
5.9.2 Site assessment	20
5.9.3 Site selection	20
5.10 Monitoring plan	21
5.11 Agreements	21
5.12 Labelling	21
6 Study site initiation	21
6.1 General	21
6.2 Prerequisites	22
6.3 Training	22
6.4 Initiation of the study site	22
7 Clinical performance study conduct	23
7.1 General	23
7.2 Responsibilities of the sponsor	23
7.3 Study site monitoring	23
7.3.1 General	23
7.3.2 Routine monitoring	23
7.3.3 Monitoring reports	24
7.4 Security and confidentiality of data	24
8 Close-out of the clinical performance study	25
8.1 Close-out activities	25
8.2 Clinical performance study report	25
8.3 Document retention	27
8.4 Suspension or premature termination of the clinical performance study	27
9 Auditing	28
Annex A (normative) Additional general requirements for certain studies	29

Annex B (normative) Clinical performance study protocol (CPSP)	32
Annex C (normative) Investigator brochure	36
Annex D (normative) Clinical performance study report	38
Annex E (normative) Communication with the ethics committee	41
Annex F (normative) Informed consent	43
Annex G (normative) Adverse event categorization	47
Annex H (informative) Good clinical performance study documentation	51
Annex I (informative) Auditing	55
Bibliography	56

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

In vitro diagnostic (IVD) medical devices are used to conduct tests outside of the human body to provide valuable information regarding a person's health or physiological status. They include tests and related devices, such as test strips and reagents, using specimens such as blood, tissue or urine, to carry out screening, diagnosis, prognosis, predictive testing, and monitoring of conditions. IVD medical devices are fundamentally different from other medical devices because they perform their function outside of the body on specimens taken from the human body. Human subjects are typically not exposed to risks with the performance testing of IVD medical devices, except for the risk associated with specimen collection procedures or when the obtained information is used for patient management. The specimens are obtained via normal body functions (e.g. urine) or through the use of invasive medical devices to allow for the specimen to be obtained (e.g. biopsy). The specimens are never reintroduced into the human body. These differences make the performance and risk characteristics of IVD medical devices different and unique from other medical devices.

Most of the studies for IVD medical devices are performed using samples resulting from the remnants of specimens taken for purposes of standard of care (leftover or archived). In these studies, there is no risk for the subjects arising from either the information provided by the IVD medical device or from the collection procedure of the specimen. However, when leftover specimens are not used, additional requirements should be considered

- when the specimens are collected specifically for the study and the specimen collection procedures present additional risk of direct harm for the subject (e.g. lumbar puncture or tissue biopsy, blood collection from neonates or critically ill patients), and/or
- when the information obtained from the IVD medical devices during the study is used to make patient management decision (i.e. interventional studies), presenting a risk of indirect harm for the subject (e.g. false negative or false positive result leading to inappropriate patient management decisions).

For the majority of IVD clinical performance studies, issues related to the use of vulnerable subjects might not arise but should be considered on a case by case basis.

Considering the reliance on specimens taken from the body and the absence of direct contact of the IVD with the patient, issues related to procedures for obtaining informed consent for IVD clinical performance studies differ from those associated with other medical devices, especially for studies with leftover or archived specimens. This document will provide guidance on the requirements for the various situations described above for IVD medical devices.

This document is intended for clinical performance studies as these studies involve specimens taken from the human body. When specimens other than leftover or archived specimens are used, there might be additional collection risks for the subject. Also in interventional studies, there might be a risk for the subject coming from the information provided by the result of the IVD under investigation.

This document is specific for IVD medical devices and therefore uses definitions and concepts that are appropriate for IVD medical devices. It is a stand-alone standard for clinical performance studies for IVD medical devices. In the situation for which there is an IVD medical device and a medical device used in an integrated system (e.g. a lancet, an IVD test strip and a glucose meter), the respective jurisdiction's regulation will define it as either an IVD medical device or a medical device and subsequently, aspects of both this document and [ISO 14155](#) might need to be considered.

Except for these situations, this document should not be read in conjunction with [ISO 14155](#), which excludes IVD medical devices from its scope.

The flowchart represented in [Figure 1](#) provides guidance on how to use this document.