

Analysis for Natural Gas and Similar Gaseous Mixtures by Gas Chromatography

Adopted as Tentative Standard, 1961 Revised and Adopted as a Standard, 1964 Revised 1972, 1986, 1989, 1990, 1995, 1999, 2000 and 2013

> **Gas Processors Association** 6526 East 60th Street Tulsa, Oklahoma 74145

DISCLAIMER

GPA publications necessarily address problems of a general nature and may be used by anyone desiring to do so. Every effort has been made by GPA to assure accuracy and reliability of the information contained in its publications. With respect to particular circumstances, local, state, and federal laws and regulations should be reviewed. It is not the intent of GPA to assume the duties of employers, manufacturers, or suppliers to warn and properly train employees, or others exposed, concerning health and safety risks or precautions.

GPA makes no representation, warranty, or guarantee in connection with this publication and hereby expressly disclaims any liability or responsibility for loss or damage resulting from its use or for the violation of any federal, state, or municipal regulation with which this publication may conflict, or for any infringement of letters of patent regarding apparatus, equipment, or method so covered.

FOREWARD

GPA 2261 provides the gas processing industry a method for determining the chemical composition of natural gas and similar gaseous mixtures using a Gas Chromatograph (GC).

The precision statements contained in this standard are based on the statistical analysis of round-robin laboratory data obtained by Section B.

This standard was developed by the cooperative efforts of many individuals from industry under the sponsorship of GPA Section B, Analysis and Test Methods.

Throughout this publication, the latest appropriate GPA Standards are referenced

"Copyright[©] 2013 by Gas Processors Association. All rights reserved. No part of this Report may be reproduced without the written consent of the Gas Processors Association."

Analysis for Natural Gas and Similar Gaseous Mixtures by Gas Chromatography

1. SCOPE

1.1 This standard covers the determination of the chemical composition of natural gas and similar gaseous mixtures within the ranges listed in Table 1, using a Gas Chromatograph (GC). The three columns represent the original Table 1, but separate the values to three distinct groups. The first group is concentrations lower than the data obtained from the round-robin project (RR-188). The second group is concentrations used in the round-robin project (RR-188). The sequations listed in the precision statement in this standard cover the range listed in the middle column, after outliers were removed. The third group is concentrations higher than the data obtained from the round-robin project (RR-188).

The precision statement in this standard utilizes equations derived from a regression of the data in RR-188 and is detailed in GPA TP-31. The precision statement criterion applies only to values listed in Section 10, Table 6.

1.2 Components sometimes associated with natural gases, i.e., helium, hydrogen sulfide, water, carbon monoxide, hydrogen and other compounds are excluded from the main body of the method. These components may be determined and made a part of the complete compositional data. Refer to Appendix A.

Component	Lower Region	Round Robin	Higher Region
Nitrogen	0.01 - 0.1	0.1 - 30	> 30
Carbon Dioxide	0.01 - 0.1	0.1 - 30	> 30
Methane	0.01 - 40	40 - 100	N / A
Ethane	0.01 - 0.1	0.1 - 10	> 10
Propane	0.01 - 0.1	0.1 - 10	> 10
Isobutane	0.01 - 0.25	0.25 - 4	> 4
n-Butane	0.01 - 0.25	0.25 - 4	> 4
Isopentane	0.01 - 0.12	0.12 - 1.5	> 1.5
n-Pentane	0.01 - 0.12	0.12 - 1.5	> 1.5
* Hexanes Plus	0.01 - 0.1	0.1 - 1.5	> 1.5
* Heptanes Plus	0.01 - 0.1	0.1 - 1.5	> 1.5

*Data from round –robin was only obtained for Hexanes Plus

Table Note: Uncertainty in the Lower region can easily be ten times greater and in the higher region two to three times greater than the center column.

NOTE 1 – Components not listed in Table 1 may be determined by procedures outlined in Appendix A or other applicable analytical procedures. Refer to Appendix A.

2. SUMMARY OF METHOD

2.1 Components to be determined in a gaseous sample are physically separated by gas chromatography and compared to calibration data obtained under identical operating conditions. A fixed volume of sample in the gaseous phase is isolated in a suitable inlet sample system and entered onto the column.

2.2 The full range analysis of a gaseous sample may require multiple runs to properly determine all components of interest. The primary run is on a partition column to determine air, methane, carbon dioxide, ethane and heavier hydrocarbons. When oxygen/argon content is critical in the unknown sample, or is suspected as a contaminant, a secondary run should be made to determine oxygen/argon and nitrogen in the air peak on the partition column. When carbon dioxide content in the unknown sample does not fall within the calibrated range on the partition column, a secondary run should be made to determine carbon dioxide content. When helium and/or hydrogen content are critical in the unknown sample, a secondary run should be made to determine helium and/or hydrogen.

2.2.1 These analyses are independent and may be made in any order, or may be made separately to obtain less than the full range analysis. The configuration can consist of a single or multiple GC's to accomplish this. Refer to Appendix A.

2.3 Response factors or response curves derived from calibration data are essential to accurately determine the composition of an unknown sample. The reference standard blend and the unknown samples must be run using identical GC operating conditions.

3. APPARATUS

3.1 *Chromatograph* - Any Gas Chromatograph may be used as long as the specifications for repeatability and reproducibility stated in *Section 10* within the round-robin test component ranges listed in *Table 1* are met or exceeded. The equipment described in this section has been proven to meet the above requirements; however other configurations including portable and online may be acceptable.

3.1.1 *Detector* - The Thermal Conductivity Detector (TCD) has proven to be a reliable and universal detector for this method.

3.1.2 *Sample Inlet System* - A gas sampling valve capable of introducing sample volumes of up to 0.500 ml may be used to introduce a fixed volume into the carrier gas stream at the head of the analyzing column. The