



Edition 1.0 2009-09

TECHNICAL REPORT



Medical device software – Part 1: Guidance on the application of ISO 14971 to medical device software





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2009 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office 3, rue de Varembé CH-1211 Geneva 20 Switzerland Email: inmail@iec.ch Web: www.iec.ch

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

Catalogue of IEC publications: <u>www.iec.ch/searchpub</u>

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

IEC Just Published: <u>www.iec.ch/online_news/justpub</u>

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

Electropedia: www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

Customer Service Centre: <u>www.iec.ch/webstore/custserv</u>

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: <u>csc@iec.ch</u> Tel.: +41 22 919 02 11 Fax: +41 22 919 03 00





Edition 1.0 2009-09

TECHNICAL REPORT



Medical device software – Part 1: Guidance on the application of ISO 14971 to medical device software

INTERNATIONAL ELECTROTECHNICAL COMMISSION



ICS 11.040.01

ISBN 2-8318-1061-9

CONTENTS

FOI	REWC)RD	4		
INTRODUCTION					
1	General				
	1.1	Scope	7		
	1.2	Normative references	7		
2	Terms and definitions				
3	General requirements for RISK MANAGEMENT				
	3.1	RISK MANAGEMENT PROCESS	8		
	3.2	Management responsibilities	.11		
	3.3	Qualification of personnel			
	3.4	RISK MANAGEMENT plan	.14		
	3.5	RISK MANAGEMENT FILE			
4	RISK	ANALYSIS	.17		
	4.1	RISK ANALYSIS PROCESS	17		
	4.2	INTENDED USE and identification of characteristics related to the SAFETY of the MEDICAL DEVICE	.18		
	4.3	Identification of HAZARDS	20		
	4.4	Estimation of the RISK(S) for each HAZARDOUS SITUATION			
5	Risk	EVALUATION	25		
6	Risk	CONTROL	26		
	6.1	RISK reduction	26		
	6.2	RISK CONTROL option analysis	26		
	6.3	Implementation of RISK CONTROL measure(s)	35		
	6.4	RESIDUAL RISK EVALUATION			
	6.5	RISK/benefit analysis			
	6.6	RISKS arising from RISK CONTROL measures			
_	6.7	Completeness of RISK CONTROL			
7		ation of overall residual risk acceptability			
8		management report			
9	Produ	uction and POST-PRODUCTION information	39		
Anr	nex A	(informative) Discussion of definitions	41		
Anr	nex B	(informative) Examples of software causes	43		
Anr	nex C	(informative) Potential software-related pitfalls	53		
Anr	nex D	(informative) Life-cycle/risk management grid	57		
Anr	nex E	(informative) SAFETY cases	60		
Bibliography					
		-			
		defined terms			
Fig	ure 1 -	- Pictorial representation of the relationship of HAZARD, sequence of events,			

HAZARDOUS SITUATION and HARM – from ISO 14971:2007 Annex E	24		
Figure 2 – FTA showing RISK CONTROL measure which prevents incorrect software			
outputs from causing HARM2	28		
Figure A.1 – Relationship between sequence of events, HARM and HAZARD4	1		

Table 1 – Requirements for documentation to be included in the RISK MANAGEMENT FILE in addition to ISO 14971:2007 requirements	17
Table A.1 – Relationship between HAZARDS, foreseeable sequences of events, HAZARDOUS SITUATIONS and the HARM that can occur	42
Table B.1 – Examples of causes by software function area	43
Table B.2 – Examples of software causes that can introduce side-effects	48
Table B.3 – Methods to facilitate assurance that RISK CONTROL methods are likely to perform as intended	52
Table C.1 – Potential software-related pitfalls to avoid	53
Table D.1 – LIFE-CYCLE/RISK MANAGEMENT grid	57

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL DEVICE SOFTWARE -

Part 1: Guidance on the application of ISO 14971 to medical device software

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committee; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 80002-1, which is a technical report, has been prepared by a joint working group of subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice, and ISO technical committee 210: Quality management and corresponding general aspects for MEDICAL DEVICES.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
62A/639A/DTR	62A/664/RVC

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table. In ISO, the technical report has been approved by 16 P-members out of 17 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this technical report the following print types are used:

- requirements and definitions: in roman type.
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS USED THROUGHOUT THIS TECHNICAL REPORT THAT HAVE BEEN DEFINED IN CLAUSE 2 AND ALSO GIVEN IN THE INDEX: IN SMALL CAPITALS.

A list of all parts of the IEC 80002 series, published under the general title *Medical device software,* can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION

Software is often an integral part of MEDICAL DEVICE technology. Establishing the SAFETY and effectiveness of a MEDICAL DEVICE containing software requires knowledge of what the software is intended to do and demonstration that the implementation of the software fulfils those intentions without causing any unacceptable RISKS.

It is important to understand that software is not itself a HAZARD, but software may contribute to HAZARDOUS SITUATIONS. Software should always be considered in a SYSTEM perspective and software RISK MANAGEMENT cannot be performed in isolation from the SYSTEM.

Complex software designs can permit complex sequences of events which may contribute to HAZARDOUS SITUATIONS. Much of the TASK of software RISK MANAGEMENT consists of identifying those sequences of events that can lead to a HAZARDOUS SITUATION and identifying points in the sequences of events at which the sequence can be interrupted, preventing HARM or reducing its probability.

Software sequences of events which contribute to HAZARDOUS SITUATIONS may fall into two categories:

- a) sequences of events representing unforeseen software responses to inputs (errors in specification of the software);
- b) sequences of events arising from incorrect coding (errors in implementation of the software).

These categories are specific to software, arising from the difficulty of correctly specifying and implementing a complex SYSTEM and the difficulty of completely verifying a complex SYSTEM.

Since it is very difficult to estimate the probability of software ANOMALIES that could contribute to HAZARDOUS SITUATIONS, and since software does not fail randomly in use due to wear and tear, the focus of software aspects of RISK ANALYSIS should be on identification of potential software functionality and ANOMALIES that could result in HAZARDOUS SITUATIONS – not on estimating probability. RISKS arising from software ANOMALIES need most often to be evaluated on the SEVERITY of the HARM alone.

RISK MANAGEMENT is always a challenge and becomes even more challenging when software is involved. The following clauses contain additional details regarding the specifics of software and provide guidance for understanding ISO 14971:2007 in a software perspective.

• Organization of the technical report

This technical report is organized to follow the structure of ISO 14971:2007 and guidance is provided for each RISK MANAGEMENT activity in relation to software.

There is some intentional REDUNDANCY in the information provided due to the iterative nature of RISK MANAGEMENT activities in the software LIFE-CYCLE.