

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-65: Particular requirements for the basic safety and essential performance
of dental intra-oral X-ray equipment**

This is a preview. [Click here to purchase the full publication.](#)



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2021 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC online collection - oc.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 18 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.



IEC 60601-2-65

Edition 1.2 2021-05
CONSOLIDATED VERSION

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-65: Particular requirements for the basic safety and essential
performance of dental intra-oral X-ray equipment**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.50

ISBN 978-2-8322-9774-2

Warning! Make sure that you obtained this publication from an authorized distributor.

® Registered trademark of

[This is a preview. Click here to purchase the full publication.](#)

REDLINE VERSION



**Medical electrical equipment –
Part 2-65: Particular requirements for the basic safety and essential performance
of dental intra-oral X-ray equipment**

This is a preview. [Click here to purchase the full publication.](#)

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
INTRODUCTION TO AMENDMENT 1	6
201.1 Scope, object and related standards.....	7
201.2 Normative references	9
201.3 Terms and definitions	10
201.4 General requirements.....	11
201.5 General requirements for testing of ME EQUIPMENT	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	12
201.7 ME EQUIPMENT identification, marking and documents	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	14
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	15
201.10 Protection against unwanted and excessive radiation HAZARDS	16
201.11 Protection against excessive temperatures and other HAZARDS	16
201.12 Accuracy of controls and instruments and protection against hazardous outputs	16
201.13 HAZARDOUS SITUATIONS and fault conditions	16
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	16
201.15 Construction of ME EQUIPMENT.....	16
201.16 ME SYSTEMS.....	16
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	17
202 Electromagnetic compatibility – Requirements and tests	17
203 Radiation protection in diagnostic X-ray equipment.....	17
Annexes	29
Annex C (informative) Guide to marking and labeling requirements for ME EQUIPMENT and ME SYSTEMS.....	30
Annex AA (informative) Particular guidance and rationale	31
Annex BB (informative) Identification of parts of dental X-RAY INTRA-ORAL SYSTEMS in relation to defined terms in this standard	38
Bibliography.....	40
Index of defined terms used in this particular standard.....	43
Figure AA.1 – AIR KERMA during IRRADIATION with direct current X-RAY GENERATOR	32
Figure AA.2 – AIR KERMA during IRRADIATION with ONE-PEAK X-RAY GENERATOR.....	34
Figure AA.3 – Waveform of long IRRADIATION TIME X-RADIATION from a ONE-PEAK X-RAY GENERATOR	35
Figure BB.1 – Structure of DENTAL INTRA-ORAL X-RAY EQUIPMENT	38
Figure BB.2 – Parts of DENTAL INTRA-ORAL X-RAY EQUIPMENT	39
Table 201.101 – List of potential ESSENTIAL PERFORMANCE to be considered by MANUFACTURER in the RISK MANAGEMENT PROCESS.....	11
Table 201.C.101 – Marking on the outside of ME EQUIPMENT or its parts	30
Table 201.C.102 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS.....	30

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-65: Particular requirements for the basic safety and essential performance of dental intra-oral X-ray equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendments has been prepared for user convenience.

IEC 60601-2-65 edition 1.2 contains the first edition (2012-09) [documents 62B/889/FDIS and 62B/897/RVD], its amendment 1 (2017-05) [documents 62B/1006/CDV and 62B/1039/RVC] and its amendment 2 (2021-05) [documents 62B/1233/FDIS and 62B/1238/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

International Standard IEC 60601-2-65 has been prepared by IEC subcommittee 62B: Diagnostic Imaging Equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This particular standard has been prepared to provide, based on IEC 60601-1:2005 and its collaterals, a complete set of BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for DENTAL INTRA-ORAL X-RAY EQUIPMENT. While the previously existing standards for such equipment were dedicated to components and subsystems, this particular standard addresses the system level of DENTAL INTRA-ORAL X-RAY EQUIPMENT. Components and their functions are addressed as far as necessary.

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of DENTAL INTRA-ORAL X-RAY EQUIPMENT.

The minimum safety requirements for DENTAL EXTRA-ORAL X-RAY EQUIPMENT are specified in a separate particular standard IEC 60601-2-63 to simplify and improve the readability

This particular standard amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for safety and essential performance*, hereinafter referred to as the general standard.

Within its specific scope, the clauses of this particular standard supersede and replace those of IEC 60601-2-7, *Medical electrical equipment – Particular requirements for the safety of high-voltage generators of diagnostic X-ray generators*. Requirements particular to DENTAL X-RAY-EQUIPMENT which were included in previous editions of the collateral standard IEC 60601-1-3 or the particular standard IEC 60601-2-28, IEC 60601-2-7 or IEC 60601-2-32 have been extracted and moved into this particular standard.

All requirements addressing integrated X-RAY TUBE ASSEMBLIES are covered by this particular standard. Therefore IEC 60601-2-28 does not apply to equipment in the scope of this International Standard.

INTRODUCTION TO AMENDMENT 1

The purpose of this first amendment to IEC 60601-2-65:2012 is to introduce changes to reference the Amendment 1 (2012) to IEC 60601-1:2005. As neither IEC 60601-2-65:2012 nor this amendment refers to specific elements of IEC 60601-1-2, the introduction of a dated reference to the latter document has been removed.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-65: Particular requirements for the basic safety and essential performance of dental intra-oral X-ray equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of DENTAL INTRA-ORAL X-RAY EQUIPMENT and its main components, hereafter also called ME EQUIPMENT.

The scope of this standard is restricted to X-RAY EQUIPMENT where the X-RAY TUBE ASSEMBLY contains the HIGH-VOLTAGE TRANSFORMER ASSEMBLY.

DENTAL EXTRA-ORAL X-RAY EQUIPMENT is excluded from the scope of this standard

NOTE 1 The X-RAY GENERATOR in DENTAL INTRA-ORAL X-RAY EQUIPMENT always comprises an X-RAY MONOBLOCK ASSEMBLY. Therefore in this particular standard the concept of X-RAY TUBE ASSEMBLY is replaced by that of X-RAY MONOBLOCK ASSEMBLY.

NOTE 2 Main components may be for instance the X-RAY MONOBLOCK ASSEMBLY and an ELECTRONIC X-RAY IMAGE RECEPTOR.

NOTE 3 Photostimulated phosphor plates and their readers (hardware and software) are excluded from the scope of this particular standard, since they have no electrical APPLIED PARTS in the PATIENT ENVIRONMENT, and are not ME EQUIPMENT.

ME EQUIPMENT and ME SYSTEMS in the scope of IEC 60601-2-63, IEC 60601-2-44, IEC 60601-2-54, IEC 60601-2-45 or IEC 60601-2-43 are excluded from the scope of this particular standard. The scope of this International Standard also excludes RADIOTHERAPY SIMULATORS and equipment for bone or tissue absorption densitometry. Excluded from the scope is also ME EQUIPMENT intended to be used for DENTAL RADIOLOGY.

Within its specific scope, the clauses of this particular standard supersede and replace those of IEC 60601-2-7, *Medical electrical equipment – Particular requirements for the safety of high-voltage generators of diagnostic X-ray generators* and of IEC 60601-2-32, *Medical electrical equipment – Particular requirements for the safety of associated equipment of X-ray equipment*.

NOTE 4 Requirements for X-RAY GENERATORS and for ASSOCIATED EQUIPMENT, which were previously specified in IEC 60601-2-7 and IEC 60601-2-32, have been included in either IEC 60601-1:2005 (Ed3) or in this particular standard. Therefore IEC 60601-2-7 and IEC 60601-2-32 are not part of the IEC 60601-1 3rd edition scheme for DENTAL INTRA-ORAL X-RAY EQUIPMENT.

All requirements addressing integrated X-RAY TUBE ASSEMBLIES are covered by this particular standard. Therefore IEC 60601-2-28 does not apply to ME EQUIPMENT in the scope of this International Standard.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*