

---

---

## Prefilled syringes —

Part 4:

### **Glass barrels for injectables and sterilized subassembled syringes ready for filling**

*Seringues préremplies —*

*Partie 4: Cylindres en verre pour produits injectables et seringues pré-  
assemblées stérilisées préremplissables*



Reference number  
ISO 11040-4:2015(E)



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>vi</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>2</b>
<b>4 General requirements</b>	<b>3</b>
4.1 Quality systems	3
4.2 Testing	3
4.3 Documentation	3
<b>5 Syringe barrel</b>	<b>4</b>
5.1 Design including dimensions	4
5.2 Functional testing of Luer connection	7
5.3 Material	7
5.4 Performance requirements	7
5.4.1 Hydrolytic resistance	7
5.4.2 Annealing quality	7
5.4.3 Lubrication of the inner surface	7
5.4.4 Flange breakage resistance	7
5.4.5 Luer cone breakage resistance	8
<b>6 Sterilized subassembled syringes ready for filling</b>	<b>8</b>
6.1 General	8
6.2 Sterility	8
6.3 Pyrogenicity/endotoxins	9
6.4 Particles	9
6.5 Additional requirements to specific components of sterilized subassembled syringes ready for filling	9
6.5.1 Barrel	9
6.5.2 Needle	10
6.5.3 Closure system	11
6.6 Closure system barrel integrity	12
<b>7 Packaging</b>	<b>12</b>
<b>8 Labelling</b>	<b>12</b>
<b>Annex A (informative) Examples of types of sterilized subassembled syringes ready for filling</b>	<b>13</b>
<b>Annex B (informative) Head designs</b>	<b>15</b>
<b>Annex C (normative) Test methods for syringe barrels</b>	<b>17</b>
<b>Annex D (informative) Sample preparation for endotoxin and particulate determination</b>	<b>23</b>
<b>Annex E (informative) Glide force test method to evaluate syringe lubrication</b>	<b>27</b>
<b>Annex F (informative) Needle penetration test</b>	<b>30</b>
<b>Annex G (normative) Test methods for closure systems</b>	<b>33</b>
<b>Annex H (informative) Dye solution tightness test</b>	<b>48</b>
<b>Bibliography</b>	<b>50</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

This third edition cancels and replaces the second edition (ISO 11040-4:2007), which has been technically revised and contains the following changes:

- Scope has been extended by adding sterilized subassembled syringes ready for filling and appropriate requirements, as well as test methods, have been included;
- general requirements have been added on quality systems, testing, and documentation;
- requirements on labelling have been revised;
- requirements on packaging have been added;
- requirements on syringes barrels have been revised by
  - adding requirements and related test methods for flange breakage resistance and Luer cone breakage resistance,
  - adding requirements on lubrication,
  - adding requirements and guidance on tolerances for Luer conical fittings, as well as on functional testing of Luer connections, and
  - deleting the clause on designation.

ISO 11040 consists of the following parts, under the general title *Prefilled syringes*:

- *Part 1: Glass cylinders for dental local anaesthetic cartridges*
- *Part 2: Plunger stoppers for dental local anaesthetic cartridges*
- *Part 3: Seals for dental local anaesthetic cartridges*

- *Part 4: Glass barrels for injectables and sterilized subassembled syringes ready for filling*
- *Part 5: Plunger stoppers for injectables*
- *Part 6: Plastics barrels for injectables*
- *Part 7: Packaging systems for sterilized subassembled syringes ready for filling*
- *Part 8: Requirements and test methods for finished prefilled syringes*

## Introduction

In the past, ampoules and injection vials were mainly used for (parenteral) injectable products. However, for the injection of the products contained in those ampoules and vials, a hypodermic syringe combined with the appropriate injection needle is also needed. This means the injectable product has to be transferred by the user into the hypodermic syringe before its final use. This procedure is not only time-consuming, but also presents a great number of possibilities for contamination.

To ensure safe use of an injectable product, prefilled syringes for single use are on the market for many years. Without a doubt, such prefilled syringes permit immediate injection of the product contained after relatively simple handling. These syringes can also be used in injectors with automated functions where further and particular requirements apply.

Based on the diameter of the prefilled syringes, appropriate components, such as rubber plungers, tip caps, needle shields, and other closure systems can also be standardized. In conjunction with the right sealing components, they offer a system for (parenteral) injectable use. The producers of filling machines can apply this part of ISO 11040 to achieve a degree of standardization in the equipment of the machines.

At the start of prefilled syringe processing by the pharmaceutical industry, syringes made of tubing glass were delivered to the pharmaceutical companies in the form of so called non-sterile “bulkware” only. The process steps washing, drying, inner lubrication, sealing the syringe with a closure system, sterilization, as well as filling and closing, were then performed in the pharmaceutical companies. Processing of “bulkware” is performed like this until today. Sterilized subassembled syringes have partially replaced non-sterile “bulkware”.

In the case of sterilized subassembled syringes ready for filling, responsibility for the aforementioned process steps relevant to the injectable product lies with the manufacturer of the primary packaging material. Following the assembly of the needle shield on syringes with a staked needle or tip caps for the Luer cone version, the subassembled syringes are placed into so called nests. The nests, in turn, are placed into a plastic tub. The syringes in the nest are protected by means of an insert liner and the tub itself is sealed by a sealing lid (which is currently and, so far, primarily achieved using a porous material). Thus, the tub properly sealed with the sealing lid represents the “sterile barrier system”. The sealed tub is then wrapped into a sealable bag and, thus, ready for sterilization which is currently and, so far, primarily performed using ethylene oxide.

In this form, the sterilized subassembled syringes ready for filling are delivered to the pharmaceutical companies in a sterile condition, where they are processed on suitable machines.

# Prefilled syringes —

## Part 4:

# Glass barrels for injectables and sterilized subassembled syringes ready for filling

## 1 Scope

This part of ISO 11040 applies to

- tubing-glass barrels (single-chamber design) for injection preparations, and
- sterilized subassembled syringes ready for filling.

It specifies materials, dimensions, quality, and performance requirements, as well as relevant test methods.

This part of ISO 11040 also specifies those components that are part of the sterilized subassembled syringe ready for filling.

Glass barrels and sterilized subassembled syringes ready for filling in accordance with this part of ISO 11040 are intended for single use only.

Components to complete the subassembled syringe, such as plunger and rod, are not specified in this part of ISO 11040.

**NOTE** Attention is drawn to applicable national or regional regulations such as Ph. Eur., USP, or JP. Where relevant, specific references to Ph. Eur., USP, and JP have been given in specific clauses or subclauses of this part of ISO 11040.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-1,<sup>1)</sup> Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements

ISO 594-2,<sup>1)</sup> Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings

ISO 720:1985, *Glass — Hydrolytic resistance of glass grains at 121 degrees C — Method of test and classification*

ISO 4802-1, *Glassware — Hydrolytic resistance of the interior surfaces of glass containers — Part 1: Determination by titration method and classification*

ISO 4802-2, *Glassware — Hydrolytic resistance of the interior surfaces of glass containers — Part 2: Determination by flame spectrometry and classification*

ISO 7864, *Sterile hypodermic needles for single use*

ISO 7886-1:1993, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use*

1) ISO 594-1 and ISO 594-2 will be replaced by ISO 80369-7 (currently in preparation by ISO/TC 210).